Ossur North America
27412 Aliso Viejo Parkway
Aliso Viejo, CA 92656, USA
Tel: +1 800 233 6263
Tel: +1 949 362 3883
Fax: +1 949 362 3888
ossurusa@ossur.com

Ossur - Braces and Support
742 Pancho Road
Camarillo, CA 93012, USA
Tel +1 805 484 2600
Fax +1 800 284 4356

Ossur - Custom Braces
19762 Pauling
Foothill Ranch, CA 92610, USA
Tel +1 949 859 4407
Fax +1 800 453 4567

Ossur Spinal
Mid Atlantic Corporate Center
1414 Metropolitan Ave.
P.O. Box 566
Thorofare, New Jersey 08086, USA
Tel: +1 856 848 0525
Fax: +1 856 845 1148
spine@ossur.com

Ossur Canada
12111 Jacobson Way
Richmond, BC
V6W 1L5, Canada
Tel: +1 604 241 8152
Fax: +1 604 241 8153

Ossur Europe
P.O. Box 120
5690 AC Son en Breugel
The Netherlands
Tel: 00800 3539 3668
Tel: +31 499 462840
Fax: +31 499 462841
saleseurope@ossur.com

Ossur Europe
Kundenservice Deutschland
Römerfeldstraße 2
50259 Pulheim, Germany
Tel: +49 (0) 2238 30 58 50
Fax: +49 (0) 2238 30 58 01
info-deutschland@ossur.com

Ossur Nordic
Salagatan 16a
P.O. Box 67
751 03 Uppsala, Sweden
Tel: +46 1818 2200
Fax: +46 1818 2218
info@ossur.com

Ossur International Markets
Grjothals 5
110 Reykjavík, Iceland
Tel: +354 515 1300
Fax: +354 515 1370
international@ossur.com

Ossur Asia Pacific
2 Redbank Road
Northmead NSW 2152
Australia
Tel +61-2-9630-9206
Fax +61-2-9630-9268
info-asiapacific@ossur.com

Equalizer® Pre-Inflated Standard Air Walker (Available in black only)

Part #	Size	Men Shoe	Women Shoe	U/M
PW0200	XSmall	2 - 4	3.5 - 5.5	1 ea
PW0400	Small	4.5 - 7	6 - 8	1 ea
PW0600	Medium	7.5 - 10.5	8.5 - 11.5	1 ea
PW0800	Large	10.5 - 12.5	11.5 - 13.5	1 ea
PW1000	XLarge	12.5+	13.5+	1 ea

Pre-Inflated Air Walker Replacement Softgoods

Part #	Size	U/M
12918	XSmall	1 ea
11058	Small	1 ea
11059	Medium	1 ea
11060	Large	1 ea
12919	XLarge	1 ea

Softgood Width Extension

Part #	Size	U/M
WKSG-1BLK	Universal	1 ea

Tibia and Calf Plates

Part #	Size	U/M
WKSP100	Universal	1 ea

Hot/Cold Therapy Insert - Gel Pad

Part #	Size	U/M
WKGEL100	Universal	1 ea

Heel Wedge

Part #	Size	U/M
B-120000004	Universal	1 ea

Ossur Head Office
Grjothals 5
110 Reykjavik, Iceland
Tel + 354 515 1300
Fax +354 515 1366
mail@ossur.com

EQUALIZER® PRE-INFLATED AIR WALKER

ÖSSUR
Life Without Limitations

www.ossur.com Copyright©Ossur IFU 0290 EN Rev.F

1

5

7

2

6

8

3

4

Equalizer® Pre-Inflated Air Walker

USES

Grade 2 and 3 Ankle Sprains, Stable Fractures, Post-Operative Use, Trauma, Rehabilitation, and Hot/Cold Therapy

CAUTION

DO NOT REMOVE WALKER UNLESS INSTRUCTED BY PHYSICIAN.

1. This product has been designed and tested based on single patient usage and is not recommended for multiple patient usage.
2. If skin irritations of any kind develop, consult medical professional immediately.
3. Be sure the patient performs circulation checks. If a loss of circulation is felt, or if the walker feels too tight, release contact closure straps and adjust to a comfortable level. If discomfort continues, the patient should contact their medical professional immediately.
4. Caution should be used if prescribed for patients with diabetes, vascular deficiencies, and neuropathies.
5. Do not remove the inner foam pads which are adhered to the plastic walker boot.

PHOTO CAPTIONS

A. Pull Tabs
B. Walker struts
C. Patented pneumatic softgood
D. Wide rocker bottom base
E. Adjustable contact closure straps
F. Foam Pads

FITTING INSTRUCTIONS

1. Remove softgood by opening contact closure straps and lifting out of brace. Do not remove pull tabs at this time.
2. Place lower leg and foot into softgood. Close softgood snugly over leg.
3. Insert foot into brace by gently spreading uprights. Do not remove pull tabs at this time.
4. Ensure struts are centered over the ankle bone (malleoli). The lower leg should be kept at a 90° angle with the foot.
5. Remove pull tabs located in between struts and softgood. This allows the softgood to be secured into position.
6. To accommodate for rubbing or spacing, additional foam pads may be placed inside the softgood, in between the softgood and plastic, or under straps for additional comfort.
7. Secure contact closure straps, starting at the toes and working upward. For best fit, secure straps in the correct order. Strap adjustments may be neccessary as foot, ankle and leg swelling patterns change.
8. When prescribed by physician, remove leg from product by releasing contact closure straps and opening softgood. Pull foot and leg forward and out, leaving softgood inside the brace. If softgood is removed, follow Steps 1 through 7 to reapply.

WASHING INSTRUCTIONS

1. This product can be washed at any time by hand using a mild detergent and cold water. Remove softgood from walker plastic before washing. Do not use bleach or strong detergent. Do not wash in washing machine or place in dryer. Air dry flat.
2. Walker plastic can be washed with a damp cloth and mild soap.

Equalizer® Pre-Inflated Air Walker is covered by one or more of the following Patent Nos. 5,078,128; D473,654. Other U.S. and foreign patents pending. ® indicates registration in U.S.A. and selected countries only.